

SL 3111 **Complete Validation Set** with bendable metal probes for large steam sterilizers



For the validation of steam sterilizer processes according to ISO 17665.

The set can individually be expanded or compiled yourself from one or more data loggers (EBI 11 or EBI 12), the appropriate interface and corresponding TÜV certified software. Talk to us!

The set contains:

- 5 x EBI 12-T441 Temperature data loggers
- 1 x EBI 12-TP 226 Temperature / pressure data logger with AL 101 silicone protection box
- EBI IF 200, 4-port Interface with USB connection and antenna
- Winlog. validation evaluation software
- Carrying case „SYSTAINER“



Type	Description	Part No.
SL 3111	Validation set for large steam sterilizers	1250-3111

Equipment for operation or process qualification

AL 3305 **DAC Adapter Set** for validation of the thermal process in DAC UNIVERSAL MK IV

With the Adapter set is an independent process qualification of a DAC Universal MK III and MK IV in combination with ebro EBI 11 data loggers possible. No modification on the customers lid necessary.



Systainer inlay is also prepared for some of adapters AL 3306 to AL 3308 (see page 54)

The set contains:

- 1 x Logger mount
- 1 x PCD test body incl. adapter for ISO, Sirona and KaVo
- Carrying case „SYSTAINER“



Type	Description	Part No.
AL 3305	DAC-Adapter Set	1248-3305